

NGS Benchmarking trial in (hemato-)oncology

Aline Antoniou

Brussels, 07/11/2017

Dept. Quality of medical laboratories

Overview

Aim

Methods

Results

- Wet and dry lab methods used
- Assessed variants and criteria
- Unexpected results:
 - Horizon artificial variants
 - Endogenous variants
- Variant description (Nomenclature)
- Clinical/biological interpretations

Conclusions

Next steps and perspectives

Aim

Assess the quality of NGS tests carried out by the participating laboratories:

Examine the ability of the validated/accredited NGS assays to accurately and reproducibly detect mutations

Overview

Aim

Methods

Results

- Wet and dry lab methods used
- Assessed variants and criteria
- Unexpected variants:
 - Horizon artificial variants
 - Endogenous variants
- Variant description (Nomenclature)
- Clinical/biological interpretations

Conclusions

Next steps and perspectives

Methods: Participants

Participation was open to all Belgian laboratories which meet the actual legal obligations:

- Licensed (pathology, clinical biology, centre of human genetics)
- Accredited or in process to be accredited for NGS testing (ISO 15189)

First Benchmarking trial focused on solid tumors


Methods: Participants

16 laboratories participated to the NGS benchmarking trial performed on solid tumors

Regions	N
Flemish region	10
Region of Brussels	4
Walloon region	2
Total	16

Laboratories	N
Pathology	10
Clinical biology	4
Center of Human Genetics	2
Total	16

Methods: Timeline



Methods: reference samples

Well-characterized genomic DNA from tumor cell lines:


→ mixture of different cell lines

Reference samples produced externally according to our custom design and validated by ddPCR (HORIZON) :

- Type and number of variants/sample
- Allele frequency between 3 and 50%
- Clinically relevant mutations

→ ComPerMed discussions

Methods: reference samples



Total of 12 analyses

Methods: request

The laboratories had to analyze samples by NGS as in routine, using their standard methods

Requests for Benchmarking trial:

- Wet lab and dry lab methods
- Variant identification and their biologic and/or clinical classification
- Raw data files (fastq, bam and vcf files)
- Clinical reports

Overview

Aim

Methods

Results

- **Wet and dry lab methods used**
- Assessed variants and criteria
- Unexpected results:
 - Horizon artificial variants
 - Endogenous variants
- Variant description (Nomenclature)
- Clinical/biological interpretations

Conclusions

Next steps and perspectives

Results: Wet and dry lab methods

Sequencer company and platform:

Company	N	Platform	N
Illumina	14	MiSeq	14
		NextSeq	3
		HiSeq	1
IonTorrent	2	PGM	2

→ Some labs have access to more than one sequencer

Results: Wet and dry lab methods

Bioinformatics pipelines:

Variant calling and annotating programs	N
Variant studio (Illumina)	6
Miseq Reporter (Illumina)	6
SeqNext (JSI)	5
NextGene (Softgenetics)	2
Sophia DDM (Sophia Genetics)	2
Torrent suite + TVC (Ion Torrent)	1
Genome Browser (Golden Helix)	1
BWA+GATK + Annovar (<i>open source</i>)	1

Results: Wet and dry lab methods

Assay Reads Characteristics:

Single/paired	N
Single end	2 (IonTorrent)
Paired end	14 (Illumina)

Minimum sequencing depth	N
1000X	7
500X	7
300X	2

Reads length	N
75bp	1
100bp	1
120bp	4
150bp	10
250bp	1

Results: Wet and dry lab methods

Which categories of somatic variants are detected by the assay?	N
SNV	16
Indels (<50bp)	16
Indels (50bp – 1kb)	0
CNV	0
Translocations	0

SNV lower limit of detection (%)	N
1%	3
2.5%	1
3%	1
4%	1
5%	10

indels < 50b Lower limit of detection (%)	N
1%	2
2.5%	2
3%	1
4%	1
5%	10

Results: Wet and dry lab methods

Which specimen types do laboratories test for somatic variant detection?	N
FFPE tissues	16
Frozen tissues	1
Fresh tissues	2
Cytological liquids	3
Blood	1
Swabs	1
Circulating tumor cell DNA	1

Do laboratories perform sequencing on tumor-normal paired specimens?	N
Yes	0
No	16

Results: Wet and dry lab methods

Required quantity of purified genomic DNA for the NGS assay	N
0-50ng	10
51-100ng	2
101-200ng	2
201-300ng	1
301-500ng	0
501-1000ng	0
>1000ng	0

- One laboratory has 2 thresholds depending on the gene panel used
- Two laboratories don't have a pre-set threshold but validate via PCR

Overview

Aim

Methods

Results

- Wet and dry lab methods used
- **Assessed variants and criteria**
- Unexpected results:
 - Horizon artificial variants
 - Endogenous variants
- Variant description (Nomenclature)
- Clinical/biological interpretations

Conclusions

Next steps and perspectives

Results: 16 assessed variants

Sample	Gene	Mutation	Proportion of laboratories that identified the mutation	Median allele frequency (%)	Standard Deviation
NGS-2017-001	BRAF	p.(Val600Glu)	16/16	13.21	0.60
NGS-2017-001	KRAS	p.(Gly13Asp)	16/16	32.94	0.64
NGS-2017-001	NRAS	p.(Gln61Lys)	16/16	21.55	1.17
NGS-2017-002	BRAF	p.(Val600Arg)	16/16	11.26	1.13
NGS-2017-002	KRAS	p.(Ala146Thr)	15/16	20.07	2.31
NGS-2017-002	NRAS	p.(Gly12Asp)	16/16	19.42	2.24
NGS-2017-003	BRAF	p.(Val600Lys)	16/16	48.50	2.95
NGS-2017-003	EGFR	p.(Glu746-Ala750del)	15/16	35.70	2.89
NGS-2017-003	EGFR	p.(Gly719Ser)	15/16	11.10	1.38
NGS-2017-003	KRAS	p.(Gly12Ala)	15/16	18.24	1.36
NGS-2017-004	BRAF	p.(Val600Met)	16/16	19.73	0.83
NGS-2017-004	EGFR	p.(Gly719Ser)	12/16	3.73	0.50
NGS-2017-004	EGFR	p.(Leu858Arg)	16/16	38.13	0.96
NGS-2017-004	EGFR	p.(Thr790Met)	16/16	38.00	1.10
NGS-2017-004	KRAS	p.(Gly12Cys)	15/16	5.16	0.42
NGS-2017-004	KRAS	p.(Gly13Asp)	16/16	29.07	0.95

→ Success criteria: Detection of all clinically relevant variants (only qualitative criteria)

Results: Success rate

Success rate (% of identified variants/16)	N
16/16 (100%)	9
15/16 (93.75%)	5
14/16 (87.50%)	2
Total success rate : 247/256 (96.48%)	16

→ In total, 9 unexpected results

Overview

Aim

Methods

Results

- Wet and dry lab methods used
- Assessed variants and criteria
- **Unexpected results:**
 - **Horizon artificial variants**
 - Endogenous variants
- Variant description (Nomenclature)
- Clinical/biological interpretations

Conclusions

Next steps and perspectives


Unexpected results

Sample	Gene	Mutation	Proportion of laboratories that identified the mutation	Median allele frequency (%)	Standard Deviation
NGS-2017-001	BRAF	p.(Val600Glu)	16/16	13.21	0.60
NGS-2017-001	KRAS	p.(Gly13Asp)	16/16	32.94	0.64
NGS-2017-001	NRAS	p.(Gln61Lys)	16/16	21.55	1.17
NGS-2017-002	BRAF	p.(Val600Arg)	16/16	11.26	1.13
NGS-2017-002	KRAS	p.(Ala146Thr)	15/16	20.07	2.31
NGS-2017-002	NRAS	p.(Gly12Asp)	16/16	19.42	2.24
NGS-2017-003	BRAF	p.(Val600Lys)	16/16	48.50	2.95
NGS-2017-003	EGFR	p.(Glu746-Ala750del)	15/16	35.70	2.89
NGS-2017-003	EGFR	p.(Gly719Ser)	15/16	11.10	1.38
NGS-2017-003	KRAS	p.(Gly12Ala)	15/16	18.24	1.36
NGS-2017-004	BRAF	p.(Val600Met)	16/16	19.73	0.83
NGS-2017-004	EGFR	p.(Gly719Ser)	12/16	3.73	0.50
NGS-2017-004	EGFR	p.(Leu858Arg)	16/16	38.13	0.96
NGS-2017-004	EGFR	p.(Thr790Met)	16/16	38.00	1.10
NGS-2017-004	KRAS	p.(Gly12Cys)	15/16	5.16	0.42
NGS-2017-004	KRAS	p.(Gly13Asp)	16/16	29.07	0.95

→ 3 Horizon artificial variants


Unexpected results: Horizon artificial variants

- Horizon artificial variant : Homologous recombination



Unexpected results: Horizon artificial variants


- Library construction can cause enrichment bias



Unexpected results: Horizon artificial variants

- Library construction can cause enrichment bias


Initial mix:



PCR
amplification



Resulting Amplicons:



→ These variants have not been evaluated for the 2 concerned laboratories.

Unexpected results: Horizon artificial variants

New success rate without the 3 Horizon artificial variants

success rate	N
100%	11
15/16 (93.75%)	4
14/16 (87.50%)	1
Total success rate : 247/253 (97.63%)	16

→ In total, 6 unexpected results in endogenous variants

Overview

Aim

Methods

Results

- Wet and dry lab methods used
- Assessed variants and criteria
- **Unexpected results:**
 - Horizon artificial variants
 - **Endogenous variants**
- Variant description (Nomenclature)
- Clinical/biological interpretations

Conclusions

Next steps and perspectives

Unexpected results: endogenous variants

Sample	Gene	Mutation	Proportion of laboratories that identified the mutation	Median allele frequency (%)	Standard Deviation
NGS-2017-001	BRAF	p.(Val600Glu)	16/16	13.21	0.60
NGS-2017-001	KRAS	p.(Gly13Asp)	16/16	32.94	0.64
NGS-2017-001	NRAS	p.(Gln61Lys)	16/16	21.55	1.17
NGS-2017-002	BRAF	p.(Val600Arg)	16/16	11.26	1.13
NGS-2017-002	KRAS	p.(Ala146Thr)	15/15	20.07	2.31
NGS-2017-002	NRAS	p.(Gly12Asp)	16/16	19.42	2.24
NGS-2017-003	BRAF	p.(Val600Lys)	16/16	48.50	2.95
NGS-2017-003	EGFR	p.(Glu746-Ala750del)	15/16	35.70	2.89
NGS-2017-003	EGFR	p.(Gly719Ser)	15/16	11.10	1.38
NGS-2017-003	KRAS	p.(Gly12Ala)	15/15	18.24	1.36
NGS-2017-004	BRAF	p.(Val600Met)	16/16	19.73	0.83
NGS-2017-004	EGFR	p.(Gly719Ser)	12/16	3.73	0.50
NGS-2017-004	EGFR	p.(Leu858Arg)	16/16	38.13	0.96
NGS-2017-004	EGFR	p.(Thr790Met)	16/16	38.00	1.10
NGS-2017-004	KRAS	p.(Gly12Cys)	15/15	5.16	0.42
NGS-2017-004	KRAS	p.(Gly13Asp)	16/16	29.07	0.95

→ 6 unexpected results concerning 3 endogenous variants

Unexpected results: endogenous variants



1. NGS-2017-003 : EGFR p.(Glu746-Ala750del)

- One laboratory reported the EGFR p.(Glu746_Ala750delinsIlePro) variant instead of the EGFR p.(Glu746-Ala750del) variant.
- Presence of the correctly identified mutation in the raw data and the vcf file.
- **Conclusion**: Transcription error from the software to the transmitted protocol.

Unexpected results: endogenous variants



2. NGS-2017-003 : EGFR p.(Gly719Ser)

- One laboratory did not identify the EGFR p.(Gly719Ser) variant with a median allele frequency of 11.10%.
- **Conclusion**: Problem of identification of this variant with the laboratory method.

Unexpected results: endogenous variants

3. NGS-2017-004 : EGFR p.(Gly719Ser)

- 4 laboratories did not report the EGFR p.(Gly719Ser) variant with a median allele frequency of 3.73%.
- 3/4 laboratories have a lower limit of detection for SNV of 5%. **Conclusion**: The allele frequency of this mutation is below their detection threshold.
- 1/4 has a lower limit of detection for SNV of 3%. **Conclusion**: Problem of identification of this variant with the laboratory method.

Overview

Aim

Methods

Results

- Wet and dry lab methods used
- Assessed variants and criteria
- Unexpected results:
 - Horizon artificial variants
 - Endogenous variants
- **Variant description (Nomenclature)**
- Clinical/biological interpretations

Conclusions

Next steps and perspectives

Results: variant description

No uniformity between laboratories for annotation of sequence variants:

In protein, cDNA and chromosomal position levels

Example:

- p.V600R
- p.Val600Arg
- **p.(Val600Arg)**
- p.Val600delinsArg
- p.V600delinsR

→ In the future, standardization required for Healthdata database (via the ComPerMed)

Results: variant description

HGVS-nomenclature: <http://varnomen.hgvs.org/>

- Recommendations for the description of sequence variants serves as an international standard

SPECIAL ARTICLE

Human Mutation

HGVS Recommendations for the Description of Sequence Variants: 2016 Update



Johan T. den Dunnen,^{1*} Raymond Dalglish,² Donna R. Maglott,³ Reece K. Hart,⁴ Marc S. Greenblatt,⁵ Jean McGowan-Jordan,⁶ Anne-Francoise Roux,⁷ Timothy Smith,⁸ Stylianos E. Antonarakis,⁹ and Peter E.M. Taschner¹⁰ on behalf of the Human Genome Variation Society (HGVS), the Human Variome Project (HVP), and the Human Genome Organisation (HUGO)

Overview

Aim

Methods

Results

- Wet and dry lab methods used
- Assessed variants and criteria
- Unexpected results:
 - Horizon artificial variants
 - Endogenous variants
- Variant description (Nomenclature)
- **Clinical/biological interpretations**

Conclusions

Next steps and perspectives

Results: clinical/biological interpretations

Example:

KRAS p.(Gly13Asp) Sample NGS-2017-001
Pred/prog in (other) tumor type
pathogene variant/COSMIC, My Cancer Genome, MD Anderson, Li et al. (2017)
pathogeen
pathogénique / impact clinique avéré
sterk klinisch significant
Ras mutatie positief. Geen indicatie voor anti-EGFR TKI behandeling bij darmtumoren (Douillard JY et al. 2013 NEJM & Berlin J. et al. 2013 NEJM & Bokemeyer C et al. JCO 2010).
Cette mutation confère une sensibilité réduite aux anticorps anti-EGFR
pathogeen/resistent aan therapie
pathogeen / De aanwezigheid van een activerende mutatie in KRAS wordt in het algemeen geassocieerd met een slechte respons op anti-EGFR monoklonale antilichaam behandeling.

→ No uniformity between laboratories for interpretation of variant

Results: clinical/biological interpretations

KRAS p.(Gly13Asp) Sample NGS-2017-001
Pred/prog in (other) tumor type
pathogene variant/COSMIC, My Cancer Genome, MD Anderson, Li et al. (2017)
pathogeen
pathogénique / impact clinique avéré
sterk klinisch significant

→ **In the future, discussion of standardization during the ComPerMed meetings**

Cette mutation confère une sensibilité réduite aux anticorps anti-EGFR
pathogeen/resistent aan therapie
pathogeen / De aanwezigheid van een activerende mutatie in KRAS wordt in het algemeen geassocieerd met een slechte respons op anti-EGFR monoklonale antilichaam behandeling.

Overview

Aim

Methods

Results

- Wet and dry lab methods used
- Assessed variants and criteria
- Unexpected results:
 - Horizon artificial variants
 - Endogenous variants
- Variant description (Nomenclature)
- Clinical/biological interpretations

Conclusions

Next steps and perspectives

Conclusions

- Numerous differences observed between NGS methods used in Belgium.
- Good global success rate (247/253): only 6 unexpected results with 3 linked to the limit of detection.
- No uniformity in sequence annotation and interpretation of the variants : ComPerMed meeting needed.

Overview

Aim

Methods

Results

- Wet and dry lab methods used
- Assessed variants and criteria
- Unexpected results:
 - Horizon artificial variants
 - Endogenous variants
- Variant description (Nomenclature)
- Clinical/biological interpretations

Conclusions

Next steps and perspectives

Next steps and perspectives

- ✓ Benchmarking trial on hematological tumors:
 - Timeline:
 - Benchmarking trial invitation: 5th July 2017
 - Sample dispatching: 9th October 2017
 - Deadline for answer: 17th November 2017
- ✓ Improvement of the methodology:
 - FFPE reference samples
 - Adaptation of the success criteria
- ✓ Development of a national EQA program

Thank you!



Aline ANTONIOU

Aline.Antoniou@wiv-isp.be

WIV-ISP

Dept. Quality of medical laboratories

www.wiv-isp.be/QML/