

**EXPERTISE AND SERVICE PROVISION  
QUALITY OF LABORATORIES  
EXTERNAL QUALITY ASSESSMENT**

**PRELIMINARY REPORT  
Proficiency Test in Veterinary Medicine  
Bovine Tuberculosis  
SURVEY 2021-5**

Expected results :

**1. Serology**

**1.1. The samples**

The panel consisted of 5 different samples with repetitions leading to a panel of 20 samples: 10 positive and 10 negative.

Sample ID	Status	repetition
PT2021BTUSERP1	Positive	5
PT2021BTUSERP2	Positive	3
PT2021BTUSERP3	Positive	2
PT2021BTUSERN1	Negative	5
PT2021BTUSERN2	Negative	5

**1.2. Randomization**

Laboratory	97504	97507	97508	97509
Sample ID				
BTUBSER21-1	PT2021BTUSERN2	PT2021BTUSERN2	PT2021BTUSERP2	PT2021BTUSERN1
BTUBSER21-2	PT2021BTUSERN2	PT2021BTUSERP1	PT2021BTUSERP1	PT2021BTUSERP2
BTUBSER21-3	PT2021BTUSERN2	PT2021BTUSERP1	PT2021BTUSERN2	PT2021BTUSERP1
BTUBSER21-4	PT2021BTUSERN1	PT2021BTUSERP1	PT2021BTUSERP1	PT2021BTUSERN2
BTUBSER21-5	PT2021BTUSERP3	PT2021BTUSERN2	PT2021BTUSERP2	PT2021BTUSERN2
BTUBSER21-6	PT2021BTUSERP1	PT2021BTUSERN1	PT2021BTUSERN2	PT2021BTUSERP2
BTUBSER21-7	PT2021BTUSERP3	PT2021BTUSERN1	PT2021BTUSERP3	PT2021BTUSERN1
BTUBSER21-8	PT2021BTUSERN1	PT2021BTUSERP2	PT2021BTUSERN1	PT2021BTUSERN1
BTUBSER21-9	PT2021BTUSERN2	PT2021BTUSERN2	PT2021BTUSERN2	PT2021BTUSERN1
BTUBSER21-10	PT2021BTUSERP1	PT2021BTUSERP3	PT2021BTUSERP1	PT2021BTUSERP3
BTUBSER21-11	PT2021BTUSERN2	PT2021BTUSERP2	PT2021BTUSERN1	PT2021BTUSERN2
BTUBSER21-12	PT2021BTUSERN1	PT2021BTUSERN2	PT2021BTUSERN1	PT2021BTUSERP3
BTUBSER21-13	PT2021BTUSERN1	PT2021BTUSERP2	PT2021BTUSERN1	PT2021BTUSERP1
BTUBSER21-14	PT2021BTUSERP2	PT2021BTUSERN2	PT2021BTUSERP3	PT2021BTUSERN2
BTUBSER21-15	PT2021BTUSERP1	PT2021BTUSERN1	PT2021BTUSERP1	PT2021BTUSERN2
BTUBSER21-16	PT2021BTUSERN1	PT2021BTUSERN1	PT2021BTUSERN1	PT2021BTUSERP2
BTUBSER21-17	PT2021BTUSERP2	PT2021BTUSERP1	PT2021BTUSERN2	PT2021BTUSERN1
BTUBSER21-18	PT2021BTUSERP2	PT2021BTUSERP3	PT2021BTUSERP1	PT2021BTUSERP1
BTUBSER21-19	PT2021BTUSERP1	PT2021BTUSERN1	PT2021BTUSERN2	PT2021BTUSERP1
BTUBSER21-20	PT2021BTUSERP1	PT2021BTUSERP1	PT2021BTUSERP2	PT2021BTUSERP1

## 2. g-interferon

### 2.1. The samples

The panel consisted of 6 different samples with repetitions leading to a panel of 20 samples: 12 positive and 8 negative.

Sample ID	Status	repetition
PT2021BTUIP1	Positive	2
PT2021BTUIP2	Positive	3
PT2021BTUIP3	Positive	5
PT2021BTUIP4	Positive	2
PT2021BTUIN1	Negative	4
PT2021BTUIN2	Negative	4

### 2.2. Randomization

Laboratory	97504	97507	97508	97509
Sample ID				
BTUBIFNg21-1	PT2021BTUIP3	PT2021BTUIN1	PT2021BTUIP2	PT2021BTUIN1
BTUBIFNg21-2	PT2021BTUIP1	PT2021BTUIP3	PT2021BTUIN1	PT2021BTUIN1
BTUBIFNg21-3	PT2021BTUIP2	PT2021BTUIN2	PT2021BTUIP3	PT2021BTUIP3
BTUBIFNg21-4	PT2021BTUIP3	PT2021BTUIP3	PT2021BTUIN1	PT2021BTUIP3
BTUBIFNg21-5	PT2021BTUIN2	PT2021BTUIP1	PT2021BTUIP3	PT2021BTUIN2
BTUBIFNg21-6	PT2021BTUIN1	PT2021BTUIP1	PT2021BTUIP3	PT2021BTUIP4
BTUBIFNg21-7	PT2021BTUIP4	PT2021BTUIP2	PT2021BTUIP1	PT2021BTUIP4
BTUBIFNg21-8	PT2021BTUIP4	PT2021BTUIN1	PT2021BTUIN1	PT2021BTUIP2
BTUBIFNg21-9	PT2021BTUIN2	PT2021BTUIP3	PT2021BTUIP2	PT2021BTUIN1
BTUBIFNg21-10	PT2021BTUIP2	PT2021BTUIP2	PT2021BTUIP4	PT2021BTUIP3
BTUBIFNg21-11	PT2021BTUIN2	PT2021BTUIP4	PT2021BTUIN2	PT2021BTUIN2
BTUBIFNg21-12	PT2021BTUIN1	PT2021BTUIN1	PT2021BTUIP4	PT2021BTUIP2
BTUBIFNg21-13	PT2021BTUIP3	PT2021BTUIP3	PT2021BTUIN2	PT2021BTUIN2
BTUBIFNg21-14	PT2021BTUIP2	PT2021BTUIN2	PT2021BTUIN2	PT2021BTUIP3
BTUBIFNg21-15	PT2021BTUIN2	PT2021BTUIN2	PT2021BTUIN2	PT2021BTUIN1
BTUBIFNg21-16	PT2021BTUIN1	PT2021BTUIN1	PT2021BTUIP2	PT2021BTUIN2
BTUBIFNg21-17	PT2021BTUIP1	PT2021BTUIN2	PT2021BTUIP1	PT2021BTUIP3
BTUBIFNg21-18	PT2021BTUIN1	PT2021BTUIP3	PT2021BTUIP3	PT2021BTUIP1
BTUBIFNg21-19	PT2021BTUIP3	PT2021BTUIP4	PT2021BTUIP3	PT2021BTUIP2
BTUBIFNg21-20	PT2021BTUIP3	PT2021BTUIP2	PT2021BTUIN1	PT2021BTUIP1

#### Authorization to release the report:



By Bernard China, scheme coordinator, on 05/08/2021.  
 Sciensano – Service Quality of laboratories  
 J. Wytsmanstreet, 14 | 1050 Bruxelles  
 Tel.: +32 2 642 53 85 E-mail : [bernard.china@ciensano.be](mailto:bernard.china@ciensano.be)

END

© Sciensano Brussels 2021.

This report may not be reproduced, published or distributed without the consent of Sciensano. The laboratories' individual results are confidential. They are not passed on by Sciensano to third parties, nor to members of the Commission, the committees of experts or the working group EQA.