

Bruxelles, le 03 Mars 2021

Qualité des Laboratoires

**RAPPORT.GLOBAL EEQ HbA1c SANG FRAIS - 2021.1A**

**Target HbA1c = 67.4 mmol/mol (Méthode de référence NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Afinion AS100 IFCC mmol/mol	67.0	1	****	****	-0.4	-0.6
Alinity C IFCC mmol/mol	65.0	1	****	****	-2.4	-3.6
Capillarys 2 flex – piercing IFCC mmol/mol	65.7	8	1.6	2.4	-1.7	-2.5
Capillarys 3 Octa IFCC mmol/mol	67.3	4	1.3	1.9	-0.2	-0.3
Capillarys 3 Tera IFCC mmol/mol	66.5	4	1.3	1.9	-0.9	-1.4
Cobas c 311/111 IFCC mmol/mol	66.9	1	****	****	-0.5	-0.8
Cobas 6000/8000 c501/502 IFCC mmol/mol	68.5	5	1.7	2.4	1.1	1.6
D-10_A1c_Program IFCC mmol/mol	67.7	3	0.6	0.9	0.2	0.3
D-100_A1c_Program IFCC mmol/mol	66.9	8	1.6	2.4	-0.6	-0.8
Menarini HA-8160 TP IFCC mmol/mol	67.0	2	0.0	0.0	-0.4	-0.6
Menarini HA-8160 VP IFCC mmol/mol	65.4	18	8.5	13.0	-2.0	-3.0
Menarini HA-8180T IFCC mmol/mol	67.0	4	0.0	0.0	-0.4	-0.6
Menarini HA-8180V IFCC mmol/mol	68.3	30	1.8	2.6	0.9	1.3
Minicap flex-piercing IFCC mmol/mol	66.3	3	0.6	0.9	-1.1	-1.6
TOSOH G11 IFCC mmol/mol	65.9	2	1.8	2.8	-1.5	-2.3
TOSOH G8 IFCC mmol/mol	68.3	33	2.1	3.0	0.9	1.3
Vantage/DCA 2000 IFCC mmol/mol	69.0	1	****	****	1.6	2.3
Variant-II_Turbo_A1c_Program IFCC mmol/mol	67.0	2	4.2	6.3	-0.4	-0.6

Votre rapport individuel est disponible sur le site  
[https://gseenet.com/CueSee\\_frame.asp?SelectMenu1=28](https://gseenet.com/CueSee_frame.asp?SelectMenu1=28)

Les coordinateurs d'enquête  
 Y. Lenga

Dr. K. Vernelen