

Bruxelles, le 11 Juin 2021

Qualité des Laboratoires

RAPPORT.GLOBAL EEQ HbA1c SANG FRAIS - 2021.3B

Target HbA1c = 64.3 mmol/mol (Méthode de référence NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Alinity C IFCC mmol/mol	66.0	2	1.4	2.1	1.7	2.7
Architect Enzymatic test C IFCC mmol/mol	64.1	1	****	****	-0.2	-0.3
Capillarys 2 flex – piercing IFCC mmol/mol	64.1	7	1.7	2.6	-0.1	-0.2
Capillarys 3 Octa IFCC mmol/mol	64.2	6	1.6	2.5	-0.1	-0.1
Capillarys 3 Tera IFCC mmol/mol	64.5	4	0.6	0.9	0.2	0.4
Cobas c 311/111 IFCC mmol/mol	64.7	1	****	****	0.4	0.7
Cobas 6000/8000 c501/502 IFCC mmol/mol	64.6	5	2.2	3.4	0.4	0.6
D-10_A1c_Program IFCC mmol/mol	66.3	3	1.2	1.7	2.1	3.2
D-100_A1c_Program IFCC mmol/mol	63.8	9	1.5	2.3	-0.4	-0.6
Menarini HA-8160 TP IFCC mmol/mol	63.0	2	2.8	4.5	-1.3	-2.0
Menarini HA-8160 VP IFCC mmol/mol	65.2	18	1.7	2.7	1.0	1.5
Menarini HA-8180T IFCC mmol/mol	65.0	4	1.4	2.2	0.7	1.1
Menarini HA-8180V IFCC mmol/mol	65.7	27	3.6	5.5	1.5	2.3
Minicap flex-piercing IFCC mmol/mol	63.7	3	0.6	0.9	-0.6	-0.9
TOSOH G11 IFCC mmol/mol	65.3	5	2.6	3.9	1.0	1.6
TOSOH G8 IFCC mmol/mol	66.0	28	1.4	2.1	1.7	2.7
Variant-II_Turbo_A1c_Program IFCC mmol/mol	66.1	2	1.3	2.0	1.8	2.8

Votre rapport individuel est disponible sur le site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=28

Les coordinateurs d'enquête
 Y. Lenga

Dr. K. Vernelen