

Bruxelles, le 09 novembre 2022

Qualité des Laboratoires

**RAPPORT GLOBAL EEQ HbA1c SANG FRAIS - 2022.6B**

**Target HbA1c = 58.5 mmol/mol (Méthode de référence NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Alinity C IFCC mmol/mol	58.0	6	1.8	3.1	-0.5	-0.9
Capillarys 2 flex – piercing IFCC mmol/mol	59.5	6	2.0	3.3	1.0	1.8
Capillarys 3 Octa IFCC mmol/mol	58.0	10	0.5	0.8	-0.5	-0.8
Capillarys 3 Tera IFCC mmol/mol	59.1	8	1.1	1.9	0.7	1.1
Cobas c 303 IFCC mmol/mol (Cobas pure)	62.0	1	****	****	3.5	6.0
Cobas c 311/111 IFCC mmol/mol	52.9	1	****	****	-5.6	-9.5
Cobas 6000/8000 c501/502 IFCC mmol/mol	58.0	2	2.8	4.9	-0.5	-0.8
D-10_A1c_Program IFCC mmol/mol	60.5	2	2.1	3.5	2.0	3.5
D-100_A1c_Program IFCC mmol/mol	57.9	9	1.0	1.8	-0.5	-0.9
Menarini HA-8160 TP IFCC mmol/mol	59.0	1	****	****	0.5	0.9
Menarini HA-8160 VP IFCC mmol/mol	58.0	1	****	****	-0.5	-0.8
Menarini HA-8180T IFCC mmol/mol	58.2	4	1.3	2.2	-0.2	-0.4
Menarini HA-8180V IFCC mmol/mol	58.5	23	2.0	3.4	0.0	0.1
Menarini HA-8190V IFCC mmol/mol	58.0	2	1.4	2.4	-0.5	-0.8
Menarini Hb NEXT IFCC mmol/mol	60.9	3	4.5	7.3	2.4	4.1
Minicap flex-piercing IFCC mmol/mol	59.0	4	0.8	1.4	0.5	0.9
TOSOH G11 IFCC mmol/mol	59.2	11	0.6	1.0	0.7	1.2
TOSOH G8 IFCC mmol/mol	60.5	31	2.9	4.7	2.1	3.5

Votre rapport individuel est disponible sur le site  
[https://gseenet.com/CueSee\\_frame.asp?SelectMenu1=28](https://gseenet.com/CueSee_frame.asp?SelectMenu1=28)

Les coordinateurs d'enquête  
 Y. Lenga

Dr. K. Vernelen