

Brussel, 12 juni 2024

**Kwaliteit van Laboratoria**

**GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.3B**

**Target HbA1c = 53.1 mmol/mol (Referentiemethode NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	59.0	3	8.7	14.8	5.9	11.1
ADAMS HA-8180V IFCC mmol/mol	51.8	22	1.1	2.2	-1.3	-2.4
ADAMS HA-8190V IFCC mmol/mol	53.5	2	0.7	1.3	0.4	0.7
Afinion AS100 IFCC mmol/mol	52.0	1	****	****	-1.1	-2.1
Alinity C IFCC mmol/mol	50.4	6	0.6	1.1	-2.7	-5.1
Capillarys 2 flex – piercing IFCC mmol/mol	53.2	4	0.9	1.8	0.1	0.2
Capillarys 3 Octa IFCC mmol/mol	53.4	12	1.2	2.2	0.3	0.6
Capillarys 3 Tera IFCC mmol/mol	53.2	5	0.4	0.8	0.1	0.2
Cobas 6000/8000 c501/502 IFCC mmol/mol	53.0	2	1.4	2.7	-0.1	-0.2
D-10_A1c_Program IFCC mmol/mol	53.7	3	2.2	4.1	0.6	1.2
D-100_A1c_Program IFCC mmol/mol	51.8	8	1.6	3.0	-1.4	-2.6
Hb NEXT IFCC mmol/mol	53.2	4	1.4	2.6	0.1	0.1
Minicap flex-piercing IFCC mmol/mol	52.3	4	1.0	1.8	-0.9	-1.6
TOSOH G11 IFCC mmol/mol	53.7	19	0.6	1.1	0.6	1.1
TOSOH G8 IFCC mmol/mol	53.9	24	1.1	2.1	0.8	1.5

Uw individueel rapport kunt u opvragen via de site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu=2](https://qseenet.com/CueSee_frame.asp?SelectMenu=2)

De enquêtecoördinatoren  
 Y. Lenga

Dr. K. Vernelen