

Brussel, 03 Maart 2021

**Kwaliteit van Laboratoria**

**GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2021.1B**

**Target HbA1c = 54.2 mmol/mol (Referentiemethode NGSP/DCCT)**

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Afinion AS100 IFCC mmol/mol	54.0	1	****	****	-0.2	-0.4
Alinity C IFCC mmol/mol	51.0	1	****	****	-3.2	-5.9
Capillarys 2 flex – piercing IFCC mmol/mol	54.1	8	1.2	2.1	-0.1	-0.1
Capillarys 3 Octa IFCC mmol/mol	54.0	5	0.7	1.3	-0.2	-0.4
Capillarys 3 Tera IFCC mmol/mol	53.0	4	0.8	1.5	-1.2	-2.2
Cobas c311/111 IFCC mmol/mol	54.0	1	****	****	-0.2	-0.4
Cobas 6000/8000 c501/502 IFCC mmol/mol	54.8	5	1.3	2.5	0.6	1.1
D-10_A1c_Program IFCC mmol/mol	54.3	3	0.6	1.1	0.1	0.2
D-100_A1c_Program IFCC mmol/mol	53.1	8	1.3	2.5	-1.1	-2.1
Menarini HA-8160 TP IFCC mmol/mol	55.0	2	1.4	2.6	0.8	1.5
Menarini HA-8160 VP IFCC mmol/mol	54.4	17	1.5	2.7	0.1	0.3
Menarini HA-8180T IFCC mmol/mol	54.5	4	0.6	1.1	0.3	0.5
Menarini HA-8180V IFCC mmol/mol	55.1	29	1.4	2.5	0.9	1.6
Minicap flex-piercing mmol/mol	53.3	3	1.2	2.2	-0.9	-1.6
TOSOH G11 IFCC mmol/mol	53.5	2	1.4	2.6	-0.7	-1.3
TOSOH G8 IFCC mmol/mol	55.4	33	1.5	2.7	1.19	2.1
Vantage/DCA 2000 IFCC mmol/mol	56.0	1	****	****	1.8	3.3
Variant-II_Turbo_A1c_Program IFCC mmol/mol	52.5	2	2.1	4.0	-1.7	-3.2

Uw individueel rapport kunt u opvragen via de site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu1=2](https://qseenet.com/CueSee_frame.asp?SelectMenu1=2)

De enquêtecoördinatoren  
 Y. Lenga

Dr. K. Vernelen