

Brussel, 03 Maart 2021

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2021.1C

Target HbA1c = 44.6 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Afinion AS100 IFCC mmol/mol	44.0	1	****	****	-0.6	-1.3
Alinity C IFCC mmol/mol	41.0	1	****	****	-3.6	-8.1
Capillarys 2 flex – piercing IFCC mmol/mol	43.9	8	1.4	3.2	-0.7	-1.5
Capillarys 3 Octa IFCC mmol/mol	43.6	5	1.7	3.8	-1.0	-2.2
Capillarys 3 Tera IFCC mmol/mol	43.8	4	0.5	1.1	-0.8	-1.9
Cobas c311/111 IFCC mmol/mol	44.3	1	****	****	-0.3	-0.6
Cobas 6000/8000 c501/502 IFCC mmol/mol	44.1	5	1.7	3.7	-0.4	-1.0
D-10_A1c_Program IFCC mmol/mol	44.7	3	1.2	2.6	0.1	0.2
D-100_A1c_Program IFCC mmol/mol	43.9	8	0.8	1.8	-0.7	-1.6
Menarini HA-8160 TP IFCC mmol/mol	44.5	2	0.7	1.6	-0.1	-0.2
Menarini HA-8160 VP IFCC mmol/mol	44.5	17	1.4	3.1	-0.1	-0.1
Menarini HA-8180T IFCC mmol/mol	41.8	4	6.5	15.6	-2.8	-6.4
Menarini HA-8180V IFCC mmol/mol	45.3	29	1.1	2.4	0.7	1.6
Minicap flex-piercing mmol/mol	43.0	3	0.0	0.0	-1.6	-3.6
TOSOH G11 IFCC mmol/mol	44.6	2	0.4	0.8	0.0	-0.1
TOSOH G8 IFCC mmol/mol	46.0	33	1.1	2.5	1.4	3.2
Vantage/DCA 2000 IFCC mmol/mol	46.0	1	****	****	1.4	3.2
Variant-II_Turbo_A1c_Program IFCC mmol/mol	42.5	2	3.5	8.3	-2.1	-4.7

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen