

Brussel, 10 juni 2022

Kwaliteit van Laboratoria

GLBAAL RAPPORT EKE HbA1c VERS BLOED - 2022.3A

Target HbA1c = 43.6 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Afinion AS100 IFCC mmol/mol	45.0	1	****	****	1.4	3.2
Alinity C IFCC mmol/mol	42.6	5	0.5	1.1	-1.0	-2.4
Capillarys 2 flex – piercing IFCC mmol/mol	43.1	8	1.6	3.8	-0.5	-1.1
Capillarys 3 Octa IFCC mmol/mol	44.1	9	2.1	4.7	0.5	1.2
Capillarys 3 Tera IFCC mmol/mol	44.7	6	1.2	2.7	1.1	2.4
Cobas c311/111 IFCC mmol/mol	43.5	1	****	****	-0.1	-0.2
Cobas 6000/8000 c501/502 IFCC mmol/mol	53.0	3	19.1	36.0*	9.4	21.5
D-10_A1c_Program IFCC mmol/mol	44.8	4	0.9	2.0	1.2	2.7
D-100_A1c_Program IFCC mmol/mol	42.1	9	1.1	2.6	-1.5	-3.5
Menarini HA-8160 TP IFCC mmol/mol	43.0	2	0.0	0.0	-0.6	-1.4
Menarini HA-8160 VP IFCC mmol/mol	44.1	4	1.9	4.4	0.5	1.2
Menarini HA-8180T IFCC mmol/mol	44.0	4	0.8	1.9	0.4	0.9
Menarini HA-8180V IFCC mmol/mol	42.2	22	1.5	3.5	-1.4	-3.2
Menarini HA-8190 IFCC mmol/mol	43.0	1	****	****	-0.6	-1.4
Menarini Hb NEXT IFCC mmol/mol	45.5	1	****	****	1.9	4.3
Minicap flex-piercing mmol/mol	43.5	4	1.3	3.0	-0.1	-0.2
TOSOH G11 IFCC mmol/mol	45.3	10	3.9	8.6	1.7	4.0
TOSOH G8 IFCC mmol/mol	43.6	32	0.9	2.1	0.0	0.0

*Individuele resultaten van methode Cobas 6000/8000 c501/502 IFCC mmol/mol (CV= 36.0%): 43.0, 41.0 en 75.0 mmol/mol.

Uw individueel rapport kunt u opvragen via de site

https://qseenet.com/CueSee_frame.asp?SelectMenu=2

De enquêtecoördinatoren

Y. Lenga

Dr. K. Vernelen