

Brussel, 10 juni 2022

Kwaliteit van Laboratoria

GBAAL RAPPORT EKE HbA1c VERS BLOED - 2022.3C

Target HbA1c = 35.2 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Afinion AS100 IFCC mmol/mol	35.0	1	****	****	-0.2	-0.5
Alinity C IFCC mmol/mol	33.6	5	0.5	1.4	-1.6	-4.5
Capillarys 2 flex – piercing IFCC mmol/mol	34.6	8	1.1	3.1	-0.6	-1.6
Capillarys 3 Octa IFCC mmol/mol	35.2	9	0.7	1.9	0.0	0.1
Capillarys 3 Tera IFCC mmol/mol	35.0	6	0.6	1.8	-0.2	-0.5
Cobas c311/111 IFCC mmol/mol	35.3	1	****	****	0.1	0.3
Cobas 6000/8000 c501/502 IFCC mmol/mol	39.3	3	11.8	30.0*	4.1	11.8
D-10_A1c_Program IFCC mmol/mol	35.6	4	1.3	3.5	0.4	1.2
D-100_A1c_Program IFCC mmol/mol	33.5	9	0.8	2.3	-1.7	-4.8
Menarini HA-8160 TP IFCC mmol/mol	34.0	2	0.0	0.0	-1.2	-3.4
Menarini HA-8160 VP IFCC mmol/mol	35.3	3	2.3	6.5	0.1	0.4
Menarini HA-8180T IFCC mmol/mol	36.3	4	0.5	1.4	1.1	3.0
Menarini HA-8180V IFCC mmol/mol	35.8	23	5.4	15.0	0.6	1.6
Menarini HA-8190 IFCC mmol/mol	34.0	1	****	****	-1.2	-3.4
Menarini Hb NEXT IFCC mmol/mol	37.0	1	****	****	1.8	5.1
Minicap flex-piercing mmol/mol	34.5	4	1.3	3.7	-0.7	-2.0
TOSOH G11 IFCC mmol/mol	40.8	10	12.0	29.0*	5.6	15.9
TOSOH G8 IFCC mmol/mol	35.7	31	0.6	1.8	0.5	1.4

*Individuele resultaten van methode Cobas 6000/8000 c501/502 IFCC mmol/mol (CV= 30.0%): 53.0, 33.0, 32.0 mmol/mol en van de methode TOSOH G11 IFCC (CV= 29.0%) 35.7, 36.3, 36.0, 33.0, 36.0, 35.4, 63.4 (2x), 33.7, 35.0 mmol/mol.

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu=2

De enquêtecoördinatoren

Y. Lenga

Dr. K. Vernelen