

Brussel, 07 november 2023

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2023.6C

Target HbA1c = 69.9 mmol/mol (Referentiemethode NGSP/DCCT)

| INSTRUMENT | RESULT (Mean) | N | SD | CV% | Bias | Bias % |
|--|---------------|----|------|------|------|--------|
| ADAMS HA-8160VP IFCC mmol/mol | 71.9 | 2 | 2.6 | 3.6 | 1.9 | 2.7 |
| ADAMS HA-8180T IFCC mmol/mol | 70.7 | 3 | 1.1 | 1.6 | 0.7 | 1.0 |
| ADAMS HA-8180V IFCC mmol/mol | 68.0 | 21 | 2.3 | 3.4 | -1.9 | -2.8 |
| ADAMS HA-8190V IFCC mmol/mol | 73.5 | 2 | 0.7 | 1.0 | 3.6 | 5.1 |
| Afinion AS100 IFCC mmol/mol | 76.0 | 1 | **** | **** | 6.1 | 8.7 |
| Alinity C IFCC mmol/mol | 66.8 | 6 | 1.8 | 2.8 | -3.2 | -4.5 |
| Capillarys 2 flex – piercing IFCC mmol/mol | 69.0 | 7 | 2.2 | 3.2 | -0.9 | -1.4 |
| Capillarys 3 Octa IFCC mmol/mol | 67.8 | 13 | 1.5 | 2.3 | -2.2 | -3.1 |
| Capillarys 3 Tera IFCC mmol/mol | 68.0 | 7 | 1.9 | 2.8 | -1.9 | -2.8 |
| Cobas 6000/8000 c501/502 IFCC mmol/mol | 70.0 | 2 | 0.0 | 0.0 | 0.1 | 0.1 |
| D-10_A1c_Program IFCC mmol/mol | 71.5 | 2 | 0.7 | 1.0 | 1.6 | 2.2 |
| D-100_A1c_Program IFCC mmol/mol | 68.7 | 8 | 1.6 | 2.3 | -1.2 | -1.7 |
| Hb NEXT IFCC mmol/mol | 72.1 | 3 | 1.2 | 1.7 | 2.1 | 3.0 |
| Minicap flex-piercing IFCC mmol/mol | 69.5 | 4 | 3.7 | 5.3 | -0.4 | -0.6 |
| TOSOH G11 IFCC mmol/mol | 70.6 | 18 | 0.9 | 1.2 | 0.7 | 1.0 |
| TOSOH G8 IFCC mmol/mol | 70.9 | 26 | 1.7 | 2.3 | 1.0 | 1.4 |

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu=2

De enquêtecoördinatoren

Y. Lenga

Dr. K. Vernelen