

Brussel, 27 januari 2025

Kwaliteit van Laboratoria

GLOBAL RAPPORT EKE HbA1c VERS BLOED - 2024.6A

Target HbA1c = 60.7 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
ADAMS HA-8180T IFCC mmol/mol	62.0	2	1.41	2.3	1.3	2.2
ADAMS HA-8180V IFCC mmol/mol	59.9	22	1.61	2.7	-0.7	-1.2
ADAMS HA-8190V IFCC mmol/mol	60.0	1	****	****	-0.7	-1.1
Afinion 2 IFCC mmol/mol	62.0	1	****	****	1.3	2.2
Afinion AS100 IFCC mmol/mol	62.0	1	****	****	1.3	2.2
Alinity C IFCC mmol/mol	58.5	1	****	****	-2.2	-3.6
Capillarys 2 flex-piercing IFCC mmol/mol	59.8	5	1.45	2.4	-0.9	-1.4
Capillarys 3 Octa IFCC mmol/mol	59.8	9	0.97	1.6	-0.9	-1.4
Capillarys 3 Tera IFCC mmol/mol	59.8	6	0.75	1.3	-0.8	-1.4
cobas b 101 IFCC mmol/mol	61.0	1	****	****	0.3	0.6
cobas c 501/502 cobas 6000 8000 IFCC	59.0	2	2.82	4.8	-1.7	-2.7
D10 A1c IFCC mmol/mol	62.3	2	2.40	3.9	1.6	2.7
D100 A1c IFCC mmol/mol	60.8	7	0.86	1.4	0.1	0.2
Hb NEXT IFCC mmol/mol	63.5	5	2.38	3.7	2.9	4.7
Minicap flex-piercing IFCC mmol/mol	57.7	4	2.21	3.8	-2.9	-4.8
TOSOH G11 IFCC mmol/mol	61.6	27	0.91	1.5	0.9	1.5
TOSOH G8 IFCC mmol/mol	61.3	16	0.76	1.2	0.6	1.1

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen/ A. Vantorre