

Brussel, 08 november 2019

**Kwaliteit van Laboratoria**

**GLBAAL RAPPORT EKE HbA1c VERS BLOED - 2019.6D**

**Target HbA1c = 54.2 mmol/mol (Referentiemethode NGSP/DCCT)**

<b>INSTRUMENT</b>	<b>RESULT (Mean)</b>	<b>N</b>	<b>SD</b>	<b>CV%</b>	<b>Bias</b>	<b>Bias %</b>
Capillarys 2 flex – piercing IFCC mmol/mol	53.9	9	0.9	1.7	-0.3	-0.6
Capillarys 3 Tera IFCC mmol/mol	54.0	2	0.0	0.0	-0.2	-0.4
Cobas c311/111 IFCC mmol/mol	55.8	1	****	****	1.6	2.9
Cobas 6000/8000 c501/502 IFCC mmol/mol	54.8	5	1.8	3.3	0.6	1.1
D-10_A1c_Program IFCC mmol/mol	55.3	4	2.1	3.7	1.0	1.9
D-100_A1c_Program IFCC mmol/mol	51.2	6	1.0	2.0	-3.1	-5.6
Menarini HA-8140 IFCC mmol/mol	54.0	1	****	****	-0.2	-0.4
Menarini HA-8160 TP IFCC mmol/mol	53.3	4	1.0	1.8	-1.0	-1.8
Menarini HA-8160 VP IFCC mmol/mol	53.4	17	1.2	2.3	-0.8	-1.5
Menarini HA-8180T IFCC mmol/mol	54.3	4	1.0	1.8	0.0	0.1
Menarini HA-8180V IFCC mmol/mol	53.9	35	1.5	2.8	-0.3	-0.5
Minicap flex-piercing mmol/mol	53.2	6	2.2	4.2	-1.0	-1.9
TOSOH G11 IFCC mmol/mol	54.9	2	0.1	0.3	0.7	1.3
TOSOH G8 IFCC mmol/mol	56.1	35	1.4	2.5	1.9	3.5
Variant-II_Turbo_A1c_Program IFCC mmol/mol	54.5	2	0.7	1.3	0.3	0.5

Uw individueel rapport kunt u opvragen via de site  
[http://www.cuesee.com/CueSee\\_frame.asp?SelectMenu1=2](http://www.cuesee.com/CueSee_frame.asp?SelectMenu1=2)

De enquêtecoördinatoren  
 Y. Lenga

Dr. K. Vernelen