

Brussel, 29 juni 2020

**Kwaliteit van Laboratoria**

**GLBAAL RAPPORT EKE HbA1c VERS BLOED - 2020.3B**

**Target HbA1c = 53.8 mmol/mol (Referentiemethode NGSP/DCCT)**

| INSTRUMENT                                 | RESULT (Mean) | N  | SD   | CV%  | Bias | Bias % |
|--|---------------|----|------|------|------|--------|
| Afinion AS100 IFCC mmol/mol                | 56.0          | 1  | **** | **** | 2.2  | 4.1    |
| Alinity C IFCC mmol/mol                    | 54.0          | 1  | **** | **** | 0.2  | 0.4    |
| Capillarys 2 flex – piercing IFCC mmol/mol | 54.2          | 9  | 1.3  | 2.4  | 0.5  | 0.8    |
| Capillarys 3 Octa IFCC mmol/mol            | 53.0          | 1  | **** | **** | -0.8 | -1.4   |
| Capillarys 3 Tera IFCC mmol/mol            | 50.3          | 3  | 2.5  | 5.0  | -3.4 | -6.4   |
| Cobas c311/111 IFCC mmol/mol               | 54.3          | 1  | **** | **** | 0.5  | 1.0    |
| Cobas 6000/8000 c501/502 IFCC mmol/mol     | 53.0          | 5  | 1.0  | 1.9  | -0.8 | -1.4   |
| D-10_A1c_Program IFCC mmol/mol             | 53.3          | 3  | 1.5  | 2.9  | -0.4 | -0.8   |
| D-100_A1c_Program IFCC mmol/mol            | 51.7          | 7  | 1.9  | 3.7  | -2.0 | -3.8   |
| Menarini HA-8140 IFCC mmol/mol             | 53.0          | 1  | **** | **** | -0.8 | -1.4   |
| Menarini HA-8160 TP IFCC mmol/mol          | 54.0          | 2  | 1.4  | 2.6  | 0.2  | 0.4    |
| Menarini HA-8160 VP IFCC mmol/mol          | 53.4          | 17 | 1.5  | 2.8  | -0.3 | -0.6   |
| Menarini HA-8180T IFCC mmol/mol            | 52.5          | 4  | 0.6  | 1.1  | -1.3 | -2.4   |
| Menarini HA-8180V IFCC mmol/mol            | 53.7          | 34 | 2.4  | 4.6  | -0.1 | -0.1   |
| Minicap flex-piercing mmol/mol             | 53.4          | 5  | 1.5  | 2.8  | -0.4 | -0.7   |
| TOSOH G11 IFCC mmol/mol                    | 52.6          | 2  | 0.6  | 1.1  | -1.2 | -2.2   |
| TOSOH G8 IFCC mmol/mol                     | 54.4          | 30 | 0.8  | 1.5  | 0.6  | 1.1    |
| Variant-II_Turbo_A1c_Program IFCC mmol/mol | 52.0          | 2  | 1.4  | 2.7  | -1.8 | -3.3   |

Uw individueel rapport kunt u opvragen via de site  
[https://qseenet.com/CueSee\\_frame.asp?SelectMenu=2](https://qseenet.com/CueSee_frame.asp?SelectMenu=2)

De enquêtecoördinatoren  
 Y. Lenga

Dr. K. Vernelen