

Brussel, 14 December 2020

Kwaliteit van Laboratoria

GBAAL RAPPORT EKE HbA1c VERS BLOED - 2020.6A

Target HbA1c = 37.3 mmol/mol (Referentiemethode NGSP/DCCT)

INSTRUMENT	RESULT (Mean)	N	SD	CV%	Bias	Bias %
Afinion AS100 IFCC mmol/mol	34.0	1	****	****	-3.3	-8.8
Alinity C IFCC mmol/mol	36.0	1	****	****	-1.3	-3.4
Capillarys 2 flex – piercing IFCC mmol/mol	36.2	9	1.8	4.9	-1.0	-2.8
Capillarys 3 Octa IFCC mmol/mol	36.0	1	****	****	-1.3	-3.4
Capillarys 3 Tera IFCC mmol/mol	36.8	4	1.5	4.1	-0.5	-1.4
Cobas c311/111 IFCC mmol/mol	37.3	1	****	****	0.0	0.1
Cobas 6000/8000 c501/502 IFCC mmol/mol	36.9	5	0.7	2.0	-0.4	-1.0
D-10_A1c_Program IFCC mmol/mol	36.0	3	2.6	7.3	-1.3	-3.4
D-100_A1c_Program IFCC mmol/mol	35.7	7	1.2	3.3	-1.5	-4.1
Menarini HA-8160 TP IFCC mmol/mol	37.7	3	1.2	3.1	0.4	1.1
Menarini HA-8160 VP IFCC mmol/mol	37.0	18	1.3	3.4	-0.3	-0.8
Menarini HA-8180T IFCC mmol/mol	37.8	4	0.5	1.3	0.5	1.3
Menarini HA-8180V IFCC mmol/mol	37.8	32	1.3	3.6	0.5	1.3
Minicap flex-piercing mmol/mol	39.8	5	8.0	20.1*	2.5	6.8
TOSOH G11 IFCC mmol/mol	37.7	3	0.7	1.8	0.4	1.1
TOSOH G8 IFCC mmol/mol	38.1	32	0.9	2.4	0.8	2.2
Variant-II_Turbo_A1c_Program IFCC mmol/mol	36.5	2	0.7	1.9	-0.8	-2.1

*Individuele resultaten van methode Minicap flex-piercing (CV = 20.1%) :54, 37, 36, 35 en 37 mmol/mol.

Uw individueel rapport kunt u opvragen via de site
https://qseenet.com/CueSee_frame.asp?SelectMenu1=2

De enquêtecoördinatoren
 Y. Lenga

Dr. K. Vernelen